EC CERTIFICATE

Full Quality Assurance System

Certificate No.: 216659-2017-CE-FRA-NA-PS Rev. 2.0

Project No.: PRJC-539917-2016-MSL-FRA Valid Until: 04 May 2024

This is to certify that the quality system of:

AIR PRODUCTS SAS

95 Avenue des Arrivaux, 38070 Saint Quentin Fallavier, France

For design, production and final product inspection/testing of:

Gases for coeliosurgery, Gases for cryosurgery

Has been assessed with respect to:

THE CONFORMITY ASSESSMENT PROCEDURE DESCRIBED IN ANNEX II EXCLUDING SECTION 4 OF COUNCIL DIRECTIVE 93/42/EEC ON MEDICAL DEVICES, AS AMENDED

and found to comply.

Further details of the product(s) and conditions for certification are given overleaf.

Place and date: Høvik, 09 December 2019



PROD 021

For: DNV GL PRESAFE AS Notified Body No.: 2460

Tone Elise Kolpus

The certificate is digitally verified by blockchain technology. For more info, see www.dnvgl.com/assurance/certificates-in-the-blockchain.html



Notice: The Certificate is subject to terms and conditions as set out in the Certification Agreement. Failure to comply may render this Certificate invalid.

MSD-CO-078-A Rev 0.0 Page 1 of 3



Certificate No.: 216659-2017-CE-FRA-NA-PS Rev. 2.0

Project No.: PRJC-539917-2016-MSL-FRA Valid Until: 04 May 2024

Jurisdiction

Application of Council Directive 93/42/EEC of 14 June 1993, adopted as "Forskrift om Medisinsk Utstyr" by the Norwegian Ministry of Health and Care Services.

Certificate history:

Revision	Description	Issue Date
0.0	Supersedes DNVGL (NB0434) certificate N° 75713-2010-CEFRA-DNV 2.0 following transfer of notified body functions to DNV Nemko Presafe AS (NB2460). Two sites removed.	2017-10-31
1.0	Recertification	2019-07-11
2.0	Scope reduction	2019-12-09

Products covered by this Certificate:

Product Description	Product Name	Class
Carbon Dioxide CO2 in cylinders of 0.6 I to 50I water volume	Gases for laparoscopic surgery	IIb
Carbon dioxide CO2, in cylinders of 2l to 50l of water volume	Gases for cryosurgery	IIa
Nitrogen N2, tank of 5l to 600l water volume	Gases for cryosurgery	IIa
Nitrous oxide N2O in cylinders of 2l to 20l water volume	Gases for cryosurgery	IIa

The complete list of devices is filed with the Notified Body

Sites covered by this certificate

Site Name	Address
AIR PRODUCTS SAS	95 avenue des Arrivaux, 38070 SAINT QUENTIN FALLAVIER, France
AIR PRODUCTS SAS	ZI de l'Epinoy, 59175 LILLE TEMPLEMARS, France



Certificate No.: 216659-2017-CE-FRA-NA-PS Rev. 2.0

Project No.: PRJC-539917-2016-MSL-FRA Valid Until: 04 May 2024

Terms and conditions

The certificate is subject to the following terms and conditions:

- Any producer (see 2001/95/EC for a precise definition) is liable for damage caused by a defect in his product(s), in accordance with directive 85/374/EEC, as amended, concerning liability of defective products.
- The certificate is only valid for the products and/or manufacturing premises listed above.
- The Manufacturer shall fulfil the obligations arising out of the quality system as approved and uphold it so that it remains adequate and efficient.
- The Manufacturer shall inform Presafe of any intended updating of the quality system and Presafe will assess the changes and decide if the certificate remains valid.
- Periodical audits will be held, in order to verify that the Manufacturer maintains and applies the quality system. Presafe reserves the right, on a spot basis or based on suspicion, to pay unannounced visits.

The following may render this Certificate invalid:

- Changes in the quality system affecting production.
- Periodical audits not held within the allowed time window.

Conformity declaration and marking of product

When meeting with the terms and conditions above, the producer may draw up an EC declaration of conformity and legally affix the CE mark followed by the Notified Body identification number of Presafe.

End of Certificate

MSD-CO-078-A Rev 0.0 Page 3 of 3